

# PIXIS Labs

Accurate. Reliable. On Time.

## Pixis Labs

12423 NE Whitaker Way

Portland, OR 97230

503-254-1794

**Job**  
**Number:** 6071930  
**Report**  
**Date:** 07/20/2016  
**ORELAP #:** OR100028  
**Project No:** 020-Willamette Primary  
School

### Cover Letter

Pat McGough  
West Linn School District  
2755 SW Borland  
West Linn, OR 97068

Dear Pat McGough,

Enclosed please find Pixis Labs analytical report for samples received as order number 6071930 on 07/19/2016. Should you have any questions about this report or any other matter, please do not hesitate to contact us. We are here to help you.

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Pixis quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be kept a maximum of 15 days from the report date unless prior arrangements have been made.

Thank you for allowing Pixis to be of service to you, we appreciate your business.

Sincerely,

Signed

Richard Reid

Project Manager

### Sample Results

<b>Sample:</b> 149-A-1	Collected: 07/19/16 07:38	Temp: 22 C	Matrix: Drinking Water
Lab ID: 114631	Received: 07/19/16 12:15	Evidence of Cooling:N	

Analyte	Result	Limit	Units	MRL	Dil.	Batch	Start/Extract	Analyzed	Notes
Method: EPA 200.8									
Copper	0.443	1.3	mg/L	0.0010	1	28863-16		07/19/16 20:20	
Lead	0.0017	0.02	mg/L	0.0002	1	28863-16		07/19/16 20:20	

<b>Sample:</b> 152-B	Collected: 07/19/16 07:34	Temp: 22 C	Matrix: Drinking Water
Lab ID: 114635	Received: 07/19/16 12:15	Evidence of Cooling:N	

Analyte	Result	Limit	Units	MRL	Dil.	Batch	Start/Extract	Analyzed	Notes
Method: EPA 200.8									
Copper	0.296	1.3	mg/L	0.0010	1	28863-17		07/19/16 20:23	
Lead	0.0017	0.02	mg/L	0.0002	1	28863-17		07/19/16 20:23	

### Laboratory Quality Control Results

EPA 200.8									
<b>QC - Initial Calibration Verif. -</b>							Batch ID: 28863-7		
Analyte	Result		Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	0.0101		0.0100	mg/L	101 %	90-110	---	---	
Lead	0.0105		0.0100	mg/L	105 %	90-110	---	---	
<b>QC - Continuing Calibration Blank - B</b>							Batch ID: 28863-10		
Analyte	Result		Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	ND			mg/L	---	---	---	---	
Lead	ND			mg/L	---	---	---	---	
<b>QC - Continuing Calibration Verif. - B</b>							Batch ID: 28863-9		
Analyte	Result		Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	0.0492		0.0500	mg/L	98 %	85-115	---	---	
Lead	0.0515		0.0500	mg/L	103 %	85-115	---	---	
<b>QC - Continuing Calibration Verif. - A</b>							Batch ID: 28863-22		
Analyte	Result		Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	0.0201		0.0200	mg/L	100 %	85-115	---	---	
Lead	0.0201		0.0200	mg/L	100 %	85-115	---	---	
<b>QC - Initial Calibration Blank -</b>							Batch ID: 28863-6		
Analyte	Result		Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	ND			mg/L	---	---	---	---	
Lead	ND			mg/L	---	---	---	---	
<b>QC - Matrix Spike - of Sample 28863 - 19</b>							Batch ID: 28863-21		
Analyte	Result	Org.Result	Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	0.303	0.291	0.0100	mg/L	116 %	70-130	---	---	
Lead	0.0121	0.0020	0.0100	mg/L	101 %	70-130	---	---	
<b>QC - Sample Duplicate - of Sample 28863 - 19</b>							Batch ID: 28863-20		
Analyte	Result	Org.Result	Spike	Units	Recovery	Limits	RPD	Limit	Notes
Copper	0.295	0.291		mg/L	---	---	1	20	
Lead	0.0020	0.0020		mg/L	---	---	0	20	

#### Abbreviations

MRL Method Reporting Limit  
ND None Detected at or above the MRL  
RPD Relative Percent Difference


Limit Maximum Contamination Level (Limit) - The water is considered safe for drinking if the analytical results are below this federal recommended action level.

***Units of Measure:***

mg/L Milligrams Per Liter

West Linn Sc  
6071930

Chain of Custody Record-Water Samples  
020-Willamette Primary School: 1403 12th St., West Linn, OR 97068



Location/Site	Date Collected	Collector's Name	Time	Date Delivered
149-A-1	7/19/2016	Steve Matthiass	7:38 AM	7/19/2016
152-B	7/19/2016	Steve Matthiass	7:34 AM	7/19/2016

Rush

22.4°C  
No Cooling

Steve Matthiass 7-19-16  
Collector's Signature Date

MRH 7-19-16 12:15  
Laboratory Signature Date